



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Applied Scale Technology
3012 Ambrose Avenue, Nashville, TN 37207

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Calibration of Weigh Scales and Analytical Balances
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:

October 1, 2003

Issue Date:

February 10, 2022

Expiration Date:

April 30, 2024

Accreditation No.:

24400

Certificate No.:

L22-116

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjlab



Certificate of Accreditation: Supplement

Applied Scale Technology

3012 Ambrose Avenue, Nashville, TN 37207

Contact Name: Johnathan Allen Phone: 615-226-6777

Accreditation is granted to the facility to perform the following calibrations:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Analytical Balance ^{FO}	2 g to 300 g	$(2 \times 10^{-4} + 1.89 \times 10^{-6}Wt) \text{ g}$	ATSM E898 Test Method Class I weights Class I balances
Bench Electronic Balance ^{FO}	5 mg to 33 kg	$(1.16 \times 10^{-2} + 1.71 \times 10^{-5}Wt) \text{ g}$	Handbook 44 Test method Class F weights Class II & III scales
Bench Electronic Balance ^{FO}	0.001 lb to 300 lb	$(1.16 \times 10^{-2} + 4.66 \times 10^{-6}Wt) \text{ lb}$	Handbook 44 Test Method Class F weights Class III scales
Platform Floor Scales ^{FO}	1 lb to 10 000 lb	$(1.152 \text{ 4} + 4.74 \times 10^{-5}Wt) \text{ lb}$	Handbook 44 Test Method Class F weights Class III scales
Fork Lift & Hopper Scales ^{FO}	50 lb to 20 000 lb	$(5.773 \text{ 4} + 2.43 \times 10^{-6}Wt) \text{ lb}$	Handbook 44 Test Method Class F weights Class III scales
Truck/Platform Crane Scale ^{FO}	20 lb to 180 000 lb	$(23.093 \text{ 9} + 5.7 \times 10^{-6}Wt) \text{ lb}$	Handbook 44 Test Method Class F weights Class III scales

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represent the smallest measurement uncertainties attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer^{FO} would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
5. The term Wt represents weight in pounds or grams (including SI multiple and submultiples units) appropriate to the uncertainty statement.